



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

OSRAM SYLVANIA, Metrology & Analytics Services

71 Cherry Hill Drive

Beverly, MA 01915

Dr. Ronald O. Daubach

Phone: 978-750-1593 Fax: 978-750-1799

E-mail: ronald.daubach@sylvania.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 100403-0

Scope Revised: 2009-09-25

OPTICAL RADIATION

NVLAP Code: 20/O02

Photometric²

Parameter

Best Uncertainty (\pm) ^{note 1}

Remarks

Luminous Flux

0.64 %

Correlated Color Temperature

6.8 K

Color

0.0005

In x and y

NVLAP Code: 20/O03

Radiometric

Spectral Irradiance Measurements²

Range in nm

Best Uncertainty (\pm) in % ^{note 1}

Remarks

250

3.7

Incandescent or Halogen

350

1.1

Incandescent or Halogen

450

0.7

Incandescent or Halogen

2009-06-29 through 2010-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 100403-0

Scope Revised: 2009-09-25

550	0.6	Incandescent or Halogen
650	0.7	Incandescent or Halogen
750	0.7	Incandescent or Halogen
850	0.6	Incandescent or Halogen
950	0.4	Incandescent or Halogen
200	6.0	Deuterium
250	2.0	Deuterium
260	2.4	Deuterium

Correlated Color Temperature

6.6 K

Total Spectral Radiant Flux

Parameter

Best Uncertainty (\pm) ^{note 1} for indicated lamp type

	FEL	9005
Level	0.7 %	0.5 %
CCT	5 K	7 K
Color (x)	0.0004	0.0005
Color (y)	0.0003	0.0004
360 nm	1.3 %	17.6 %
400 nm	1.1 %	1.4 %
450 nm	0.8 %	1.0 %
550 nm	0.8 %	0.7 %
650 nm	0.6 %	0.6 %
750 nm	0.5 %	0.5 %
830 nm	0.6 %	0.6 %

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. All photometric and radiometric calibration services are for internal calibration and are not available outside of OSRAM Sylvania.

2009-06-29 through 2010-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology