



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Henry Troemner, LLC
 201 Wolf Drive
 P.O. Box 87
 Thorofare, NJ 08086-0087
 Mr. Wilbert D. Abele
 Phone: 856-686-1600 Fax: 856-686-1601
 E-mail: wabele@troemner.com
 URL: http://www.troemner.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0

Scope Revised: 2009-10-14

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

DIMENSIONAL

NVLAP Code: 20/D05
 Length Indicators- Comparison to Gage Blocks

<i>Range in Inches</i>	<i>Best Uncertainty (±) in uinches ^{note 1}</i>
0.05 to 1	68
2	94
3	138
4	174
6	177
10	211

ELECTROMAGNETICS – DC LOW FREQUENCY

NVLAP Code: 20/E02
 AC Current

<i>Range</i>	<i>Best Uncertainty (±) in ppm ^{note 1}</i>					
	<i>Frequency in Hertz</i>					
	<i>10 Hz</i>	<i>60 Hz</i>	<i>1 kHz</i>	<i>5 kHz</i>	<i>10 kHz</i>	<i>30 kHz</i>
100 µA	950	375	375	375	375	

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0

Scope Revised: 2009-10-14

1 mA	100	50	75	120	1250
10 mA	75	75	125	125	1300
100 mA	70	70	90	70	1250
1 A	400	400	400	950	
10 A	150	150	1400		

NVLAP Code: 20/E05
DC Current – Generated

Range

- 100 μ A
- 1 mA
- 10 mA
- 100 mA
- 1 A
- 10 A

Best Uncertainty (\pm) in ppm *note 1*

- 20
- 20
- 20
- 20
- 100
- 150

NVLAP Code: 20/E05
DC Resistance

Range in ohms

- 1
- 10
- 100
- 1 k
- 10 k
- 100 k
- 1 M
- 10 M
- 100 M
- 1 G

Best Uncertainty (\pm) in ppm *note 1*

- 125
- 20
- 15
- 10
- 10
- 10
- 10
- 20
- 125
- 1110

NVLAP Code: 20/E06
DC Voltage

Range in volts

- 10 mV
- 100 mV

Best Uncertainty (\pm) in ppm *note 1*

- 140
- 16

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0

Scope Revised: 2009-10-14

1 V	7.2
10 V	8.0
100 V	7.2
1000 V	6.5

NVLAP Code: 20/E09
LF AC Voltage

Best Uncertainty (\pm) in ppm ^{note 1}
Frequency in Hertz

Range	10	60	1 k	5 k	10 k	30 k
10 mV			410			
100 mV	60	60	75	60	75	125
1 V	75	65	55	55	70	150
10 V	150	125	125	125	150	225
100 V		110	110	115	120	175
1000 V		100	50	50	125	

MECHANICAL

NVLAP Code: 20/M08
Mass

Range	Best Uncertainty (\pm)	Remarks
50 kg	5.6 mg	Echelon I
30 kg	3.6 mg	Echelon I
25 kg	3.1 mg	Echelon I
20 kg	2.0 mg	Echelon I
10 kg	0.61 mg	Echelon I
5 kg	0.33 mg	Echelon I
3 kg	0.16 mg	Echelon I
2 kg	0.10 mg	Echelon I
1 kg	0.043 mg	Echelon I
500 g	0.027 mg	Echelon I
300 g	0.018 mg	Echelon I
200 g	0.012 mg	Echelon I

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0
Scope Revised: 2009-10-14

100 g	0.0069 mg	Echelon I
50 g	0.0055mg	Echelon I
30 g	0.004 mg	Echelon I
20 g	0.0029 mg	Echelon I
10 g	0.0019 mg	Echelon I
5 g	0.0012 mg	Echelon I
3 g	0.00092 mg	Echelon I
2 g	0.00072 mg	Echelon I
1 g	0.00055 mg	Echelon I
500 mg	0.0005 mg	Echelon I
300 mg	0.00051 mg	Echelon I
200 mg	0.00046 mg	Echelon I
100 mg	0.00043 mg	Echelon I
50 mg	0.00039 mg	Echelon I
30 mg	0.00040 mg	Echelon I
20 mg	0.00036 mg	Echelon I
10 mg	0.00033 mg	Echelon I
5 mg	0.00036 mg	Echelon I
3 mg	0.00039 mg	Echelon I
2 mg	0.00035 mg	Echelon I
1 mg	0.00029 mg	Echelon I
500 kg	1.2 g	Echelon II
250 kg	0.53 g	Echelon II
200 kg	0.46 g	Echelon II
100 kg	0.44 g	Echelon II
50 kg	94 mg	Echelon II
30 kg	45 mg	Echelon II
25 kg	28 mg	Echelon II
20 kg	24 mg	Echelon II
10 kg	12 mg	Echelon II
5 kg	10 mg	Echelon II
3 kg	10 mg	Echelon II
2 kg	6.6 mg	Echelon II

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0
Scope Revised: 2009-10-14

1 kg	0.40 mg	Echelon II
500 g	0.30 mg	Echelon II
300 g	0.23 mg	Echelon II
200 g	0.18 mg	Echelon II
100 g	0.10 mg	Echelon II
50 g	0.09 mg	Echelon II
30 g	0.069 mg	Echelon II
20 g	0.052 mg	Echelon II
10 g	0.042 mg	Echelon II
5 g	0.014 mg	Echelon II
3 g	0.0080 mg	Echelon II
2 g	0.0080 mg	Echelon II
1 g	0.005 mg	Echelon II
500 mg	0.005 mg	Echelon II
300 mg	0.004 mg	Echelon II
200 mg	0.004 mg	Echelon II
100 mg	0.004 mg	Echelon II
50 mg	0.0040 mg	Echelon II
30 mg	0.0035 mg	Echelon II
20 mg	0.0035 mg	Echelon II
10 mg	0.0035 mg	Echelon II
5 mg	0.0035 mg	Echelon II
3 mg	0.0034 mg	Echelon II
2 mg	0.0035 mg	Echelon II
1 mg	0.0036 mg	Echelon II
1000 lb	1.8 g	Echelon II
500 lb	0.65 g	Echelon II
200 lb	0.53 g	Echelon II
100 lb	0.16 g	Echelon II
50 lb	68 mg	Echelon II
25 lb	16 mg	Echelon II
20 lb	6 mg	Echelon II
10 lb	4 mg	Echelon II

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0
Scope Revised: 2009-10-14

5 lb	2 mg	Echelon II
3 lb	2 mg	Echelon II
2 lb	0.47 mg	Echelon II
1 lb	0.38 mg	Echelon II
0.5 lb	0.28 mg	Echelon II
0.3 lb	0.107 mg	Echelon II
0.2 lb	0.057 mg	Echelon II
0.1 lb	0.039 mg	Echelon II
0.05 lb	0.039 mg	Echelon II
0.03 lb	0.024 mg	Echelon II
0.02 lb	0.023 mg	Echelon II
0.01 lb	0.022 mg	Echelon II
0.005 lb	0.014 mg	Echelon II
0.003 lb	0.013 mg	Echelon II
0.002 lb	0.011 mg	Echelon II
0.001 lb	0.011 mg	Echelon II
1000 kg	10.3 g	Echelon III
500 kg	5.0 g	Echelon III
200 kg	3.3 g	Echelon III
100 kg	1.6 g	Echelon III
50 kg	0.087 g	Echelon III
30 kg	0.072 g	Echelon III
25 kg	0.066 g	Echelon III
20 kg	0.057 g	Echelon III
10 kg	0.024 g	Echelon III
5 kg	18 mg	Echelon III
3 kg	17 mg	Echelon III
2 kg	12 mg	Echelon III
1 kg	10 mg	Echelon III
500 g	10 mg	Echelon III
300 g	10 mg	Echelon III
3000 lb	17 g	Echelon III
2500 lb	14 g	Echelon III

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0
Scope Revised: 2009-10-14

2000 lb	10 g	Echelon III
1000 lb	5.2 g	Echelon III
500 lb	3.8 g	Echelon III
100 lb	0.088 g	Echelon III
50 lb	0.054 g	Echelon III
30 lb	0.046 g	Echelon III
25 lb	0.035 g	Echelon III
20 lb	0.029 g	Echelon III
10 lb	0.018 g	Echelon III
5 lb	11 mg	Echelon III
3 lb	10 mg	Echelon III
2 lb	10 mg	Echelon III
1 lb	10 mg	Echelon III
0.5 lb	10 mg	Echelon III

Magnetic Susceptibility (χ) of mass standards or materials used to manufacture mass standards

<i>Mass Range</i>	<i>Best Uncertainty (\pm) in χ ^{notes 1, 2}</i>	<i>Remarks</i>
10 kg	0.000464	
5 kg	0.000375	
3 kg	0.000339	
2 kg	0.000316	
1 kg	0.000349	
500 g	0.000438	
300 g	0.000488	
200 g	0.000515	
100 g	0.000567	
50 g	0.000628	
30 g	0.000654	
20 g	0.000665	
10 g	0.000462	
5 g	0.000682	
3 g	0.000684	

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0
Scope Revised: 2009-10-14

2 g 0.000685
1 g 0.000986

NVLAP Code: 20/M12
Volume - Pipettes

Gravimetric method

<i>Test Volume in μl</i> ^{note 4}	<i>Best Uncertainty (\pm) in μl</i> ^{notes 1, 3}	<i>Remarks</i>
0.1	0.03	
0.2	0.03	
0.5	0.03	
1.0	0.038	
2.5	0.047	
5.0	0.056	
10	0.12	
25	0.12	
50	0.26	
100	0.41	
200	0.44	
500	1.37	
1000	2.05	
2500	5.25	
5000	10	
10 000	20	

Spectrophotometric method

<i>Test Volume in μl</i> ^{note 4}	<i>Best Uncertainty (\pm) in μl</i> ^{notes 1, 3}	<i>Remarks</i>
0.1	0.031	
0.2	0.026	
0.5	0.050	
1.0	0.078	
2.5	0.119	
5	0.251	

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0
Scope Revised: 2009-10-14

10	0.339
25	0.39
50	1.87
100	2.60
500	13.12
1000	16.76
2500	23.08

Density in the Range of 2.69 g/cm³ to 8.5 g/cm³

<i>Mass Range</i>	<i>Best Uncertainty (±) ^{note 1}</i>	<i>Remarks</i>
5 kg	0.0056 g/cm ³	
3 kg	0.0041 g/cm ³	
2 kg	0.0034 g/cm ³	
1 kg	0.0014 g/cm ³	
500 g	0.0064 g/cm ³	
300 g	0.0075 g/cm ³	
200 g	0.0053 g/cm ³	
100 g	0.0031 g/cm ³	
50 g	0.0030 g/cm ³	
30 g	0.0067 g/cm ³	
20 g	0.0029 g/cm ³	
10 g	0.0065 g/cm ³	
5 g	0.0170 g/cm ³	
3 g	0.0148 g/cm ³	
2 g	0.0127 g/cm ³	
1 g	0.0156 g/cm ³	

THERMODYNAMIC

NVLAP Code: 20/T02
Humidity Generation

<i>Range in %</i>	<i>Best Uncertainty (±) in % ^{note 1}</i>	<i>Remarks</i>
10 to 95	0.7	Relative Humidity

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0

Scope Revised: 2009-10-14

NVLAP Code: 20/T05

Pressure

Range

Best Uncertainty (\pm) ^{note 1}

Pneumatic Pressure using Piston Gauge (absolute mode) – Direct Pressure Comparison

2.0 psia to 10.0 psia (13.8 kPa to 68.9 kPa)	30.5 ppm + 0.2 Pa
10.0 psia to 50.0 psia (68.9 kPa to 345.0 kPa)	19.3 ppm + 0.2 Pa
50.0 psia to 100.0 psia (345.0 kPa to 689.500 kPa)	46.1 ppm + 2.0 Pa
100 psia to 1000 psia (689.5 kPa to 6895.0 kPa)	22.5 ppm + 2.0 Pa

Pneumatic Pressure using Piston Gauge (Gauge Mode) – Direct Pressure Comparison

2.0 psig to 10.0 psig (13.8 kPa to 68.9 kPa)	30.5 ppm + 0.2 Pa
10.0 psig to 50.0 psig (68.9 kPa to 345.0 kPa)	19.3 ppm + 0.2 Pa
50.0 psig to 100.0 psig (345.0 kPa to 689.500 kPa)	46.1 ppm + 2.0 Pa
100 psig to 1000 psig (689.5 kPa to 6895.0 kPa)	22.5 ppm + 2.0 Pa

Pneumatic Effective Area using Piston Gauge- Direct Pressure Comparison

2 psi to 50 psi (13.8 kPa to 344.7 kPa)	16.4 ppm + 0.2 Pa
100 psi to 1000 psi (689.5 kPa to 6895 kPa)	25.8 ppm + 2.0 Pa

Pneumatic Pressure using Precision Transducer

100 to 200 psi (1000 psi range on PPC3)	273 ppm + 2.0 Pa
200 to 500 psi (1000 psi range on PPC3)	145 ppm + 2.0 Pa
500 to 1000 psi (1000 psi range on PPC3)	70 ppm + 2.0 Pa
1 to 10 psi (100 psi range on PPC3)	1381 ppm + 0.2 Pa
10 to 50 psi (100 psi range on PPC3)	126 ppm + 0.2 Pa
50 to 100 psi (100 psi range on PPC3)	58 ppm + 2.0 Pa

NVLAP Code: 20/T07

Digital Thermometry Indicators

Readout devices that actually measure resistance

Range

Best Uncertainty (\pm) ^{note 1}

Remarks

10 Ω	8 ppm	Resistance Function
25 Ω	2 ppm	Resistance Function

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0

Scope Revised: 2009-10-14

100 Ω	2 ppm	Resistance Function
300 Ω	2 ppm	Resistance Function
10 000 Ω	10 ppm	Resistance Function
1 Ω to 400 Ω	8 ppm	Resistance Function
400 Ω to 10 000 Ω	26 ppm	Resistance Function

Digital Thermometer with PRT System

-196 °C	8 mk	Comparison Method
-40 °C to 0.0 °C	8 mk	Comparison Method
0.05 °C to 95 °C	8 mk	Comparison Method
90 °C to 156 °C	10 mk	Comparison Method
156 °C to 250 °C	16 mk	Comparison Method
250 °C to 420 °C	20 mk	Comparison Method
420 °C to 550 °C	21 mk	Comparison Method

NVLAP Code: 20/T07
Resistance Thermometry

Range in °C	Best Uncertainty (±) in mk ^{note 1}	Remarks
-196	8	Comparison Method
-40 to 0.0	8	Comparison Method
-0.5 to 95	8	Comparison Method
95 to 156	10	Comparison Method
156 to 250	16	Comparison Method
250 to 420	20	Comparison Method
420 to 660	21	Comparison Method
-40 to 0	8	Thermistors
0 to 95	8	Thermistors
95 to 150	10	Thermistors

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 105013-0

Scope Revised: 2009-10-14

NVLAP Code: 20/T08
Thermocouple Indicator

<i>Range in mV</i>	<i>Best Uncertainty (±) in ppm ^{note 1}</i>
-100	42
-10	231
-1	1922
1	1778
10	222
100	41

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. Approximate value. Actual value determined by the test statistics.
3. Uncertainties at specified test volumes may be greater depending on the range of the unit under test.
4. It is recommended that adjustable volume pipettes not be used below 10 % of capacity.

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology