



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

IPS Corporation

1878-1, Harumiya Ono, Tatsuno-machi,
Kamiina-gun, Nagano 399-0601
JAPAN

Mr. Nobuyuki Kuwazawa

Phone: +81-266-44-5200 Fax: +81-266-44-5300

E-Mail: qa@ips-emc.co.jp

URL: <http://www.ips-emc.co.jp>

CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

ELECTROMAGNETICS - RF MICROWAVE, DC/LF, FREQUENCY

(Capabilities listed by instrument type and/or method)

NVLAP Code: 20/R08

Microwave Antenna Parameters

<i>Range</i>	<i>Best Uncertainty</i> (\pm) ^{note 1}	<i>Remarks</i>
Dipole Antenna (such as the VHA9103/UHA9105)		
Reference Antenna Method (substitution method)		
30 MHz to 80 MHz (Tuned 80 MHz)	0.6 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) EMI Receiver used
30 MHz to 300 MHz	0.6 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) EMI Receiver used
300 MHz to 1000 MHz	0.8 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) EMI Receiver used

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

Biconical Antenna (such as the BBA9106)

Reference Antenna Method (substitution method)

30 MHz to 300 MHz	0.8 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) EMI Receiver used
30 MHz to 300 MHz	0.7 dB	Horizontal Antenna Factor (D = 3 m, H = 2 m) EMI Receiver used
30 MHz to 300 MHz	0.8 dB	Horizontal Antenna Factor (D = 3 m, H = 1 m) EMI Receiver used
30 MHz to 300 MHz	0.8 dB	Vertical Antenna Factor (D = 3 m, H = 1.5 m) EMI Receiver used
30 MHz to 300 MHz	1.0 dB	Vertical Antenna Factor (D = 3 m, H = 1 m) EMI Receiver used
30 MHz to 300 MHz	0.7 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) Network Analyzer used
30 MHz to 300 MHz	0.6 dB	Horizontal Antenna Factor (D = 3 m, H = 2 m) Network Analyzer used
30 MHz to 300 MHz	0.7 dB	Horizontal Antenna Factor (D = 3 m, H = 1 m) Network Analyzer used
30 MHz to 300 MHz	0.8 dB	Vertical Antenna Factor (D = 3 m, H = 1.5 m) Network Analyzer used
30 MHz to 300 MHz	0.9 dB	Vertical Antenna Factor (D = 3 m, H = 1 m) Network Analyzer used

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

Log-Periodic Antenna (such as the USLP9143/UHALP9108A)

Reference Antenna Method (substitution method)

300 MHz to 1000 MHz	1.2 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) EMI Receiver used
300 MHz to 1000 MHz	1.1 dB	Horizontal Antenna Factor (D = 3 m, H = 2 m) EMI Receiver used
300 MHz to 1000 MHz	1.1 dB	Horizontal Antenna Factor (D = 3 m, H = 1 m) EMI Receiver used
300 MHz to 1000 MHz	1.1 dB	Vertical Antenna Factor (D = 3 m, H = 1.5 m) EMI Receiver used
300 MHz to 1000 MHz	1.2 dB	Vertical Antenna Factor (D = 3 m, H = 1 m) EMI Receiver used
300 MHz to 1000 MHz	1.1 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) Network Analyzer used
300 MHz to 1000 MHz	1.1 dB	Horizontal Antenna Factor (D = 3 m, H = 2 m) Network Analyzer used
300 MHz to 1000 MHz	1.1 dB	Horizontal Antenna Factor (D = 3 m, H = 1 m) Network Analyzer used
300 MHz to 1000 MHz	1.1 dB	Vertical Antenna Factor (D = 3 m, H = 1.5 m) Network Analyzer used
300 MHz to 1000 MHz	1.2 dB	Vertical Antenna Factor (D = 3 m, H = 1 m) Network Analyzer used

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

Bi-log Antenna (such as the CBL6112B)

Reference Antenna Method (substitution method)

30 MHz to 1000 MHz	1.3 dB	Horizontal Antenna Factor (D = 10 m, H = 2 m) EMI Receiver used
30 MHz to 1000 MHz	1.3 dB	(D = 10 m, H = 2 m) Network Analyzer used
30 MHz to 1000 MHz	1.3 dB	(D = 3 m, H = 2 m) Network Analyzer used

Biconical Antenna
SAE ARP958

25 to 300 MHz	0.8 dB	Horizontal Antenna Factor (D = 1 m, H = 3 m) Network Analyzer Used
25 to 300 MHz	0.6 db	Vertical Antenna Factor (D = 1 m, H = 3 m)

Log-Periodic Antenna
SAE ARP958

150 to 300 MHz	1.1 dB	Horizontal Antenna Factor (D = 1 m, H = 3 m) Network Analyzer Used
300 to 1000 MHz	0.4 dB	Horizontal Antenna Factor (D = 1 m, H = 3 m) Network Analyzer Used
1000 to 1800 MHz	0.7 dB	Horizontal Antenna Factor (D = 1 m, H = 3 m) Network Analyzer Used
150 to 300 MHz	0.6 dB	Vertical Antenna Factor (D = 1 m, H = 3 m) Network Analyzer Used

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

300 to 1000 MHz	0.2 dB	Vertical Antenna Factor (D = 1 m, H = 3 m)
1000 to 1800 MHz	0.5 dB	Network Analyzer Used Vertical Antenna Factor (D = 1 m, H = 3 m) Network Analyzer Used

NVLAP Codes: 20/E06, 20/E09, 20/R13

LISN *note 2*

9 kHz to 108 MHz	2.0 %	Impedance
9 kHz to 108 MHz	0.2 dB	Insertion Loss
AC 100 V to 240 V with 50 Hz	0.17 %	Series Voltage Drop
DC 100 V to 28 V	0.15 %	Series Voltage Drop

CDN *note 2*

100 kHz to 230 MHz	1.35 %	Impedance
100 kHz to 230 MHz	0.2 dB	Insertion Loss

NVLAP Codes: 20/E05, 20/E06, 20/E09, 20/F04

ESD Simulators

7.5 A to 60A (2 kV to 15 kV)	2.7 %	Peak current Agilent 54845A, WavePro 725Zi
0.7 ns to 1 ns (2 V to 1.5 kV)	6.2 %	Rise time Agilent 54845A, WavePro 725Zi
4 A to 30 A (2 kV to 15 kV)	3.3 %	Discharge current at 30 ns Agilent 54845A, WavePro 725Zi
2 A to 15 A (2 kV to 15 kV)	5.2 %	Discharge current at 60 ns Agilent 54845A, WavePro 725Zi
0.5 to 30 kV	1.2 %	DC High Voltage, DHM-40/10

EFT/Burst *note 2*

0.25 kV to 4 kV	2.8 %	Peak voltage Agilent 54845A, WavePro 725Zi
-----------------	-------	---

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

3.5 ns to 6.5 ns (0.25 kV to 4 kV)	1.1 %	Rise time Agilent 54845A, WavePro 725Zi
35 ns to 65 ns (0.25 kV to 4 kV)	1.0 %	Impulse duration Agilent 54845A, WavePro 725Zi
2.5 kHz to 5 kHz (0.25 kV to 4 kV)	1.2 %	Repetition frequency, 54622A
12 ms to 18 ms (0.25 kV to 4 kV)	1.0 %	Burst duration, 54622A
240 ms to 360 ms (0.25 kV to 4 kV)	1.0 %	Burst period, 54622A
Surge Generator ^{note 2} (IEC 61000-4-5)		
0.1 to 4 kV	3.0 %	Peak Voltage, Open circuit waveform Agilent 54845A, WavePro 725Zi
0.7 to 1.6 μs	3.3 %	Rise Time, open circuit waveform Agilent 54845A, WavePro 725Zi
15 to 60 μs	3.7 %	Half value duration time Agilent 54845A, WavePro 725Zi
0.1 to 2 kA	2.7 %	Peak Current, short circuit waveform Agilent 54845A, WavePro 725Zi
1.4 to 9.6 μs	3.2 %	Rise Time, short circuit waveform Agilent 54845A, WavePro 725Zi
12 to 33 μs	2.3 %	Half value duration time Agilent 54845A, WavePro 725Zi
NVLAP Code: 20/R13		
Directional Coupler ^{note 2}		
9 kHz to 6 GHz, 0 dB to 60 dB	0.20 dB	Coupling Factor
9 kHz to 6 GHz, 0 dB to 70 dB	0.44 dB	Coupling Factor
9 kHz to 6 GHz, 0 dB to 80 dB	0.95 dB	Coupling Factor
9 kHz to 6 GHz	0.20 dB	Insertion Loss
EM Clamp ^{note 2}		
0.1 MHz to 230 MHz	0.4 dB	Insertion Loss

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

150 ohm to 50 ohm Adaptor ^{note 2}

0.1 MHz to 230 MHz	0.3 dB	Insertion Loss
--------------------	--------	----------------

Current Probe/Current Injection Probe ^{note 2}

10 kHz to 500 MHz	0.5 dB	Insertion Loss
-------------------	--------	----------------

10 kHz to 500 MHz	0.5 dB ohm	Transfer Impedance
-------------------	------------	--------------------

Hi-Impedance Probe ^{note 2}

9 kHz to 30 kHz	0.3 dB	Voltage Division Factor
-----------------	--------	-------------------------

Impedance & VSWR Measure ^{note 2}

9 kHz to 500 MHz	2.0 %	4395A + 87512A
------------------	-------	----------------

100 kHz to 1.8 GHz	2.0 %	4396B
--------------------	-------	-------

300 kHz to 3 GHz	1.2 %	E5071B
------------------	-------	--------

3 GHz to 6 GHz	2.2 %	E5071B
----------------	-------	--------

RF Insertion Loss/Gain Measure ^{note 2}

9 kHz to 500 MHz, 0 dB to 60 dB	0.20 dB	4395A + 87512A
---------------------------------	---------	----------------

100 kHz to 1.8 GHz, 0 dB to 60 dB	0.15 dB	4396B
-----------------------------------	---------	-------

300 kHz to 3 GHz, 0 dB to 60 dB	0.10 dB	E5071B
---------------------------------	---------	--------

3 GHz to 6 GHz, 0 dB to 60 dB	0.13 dB	E5071B
-------------------------------	---------	--------

NVLAP Code: 20/E09

AC Voltage Measure (0.1 V to 750 V) ^{note 2}

3 Hz to 5 Hz	3.1 %	34401A
--------------	-------	--------

5 Hz to 10 Hz	0.46 %	34401A
---------------	--------	--------

10 Hz to 20 kHz	0.12 %	34401A
-----------------	--------	--------

20 kHz to 50 kHz	0.18 %	34401A
------------------	--------	--------

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200012-0

50 kHz to 100 kHz	0.79 %	34401A
100 kHz to 300 kHz	5.2 %	34401A

NVLAP Code: 20/E06
DC Voltage Measure ^{note 2}

0.01 V to 1000 V	0.11 %	34401A
------------------	--------	--------

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
 2. Items available for on-site service.

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology