



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

<p>Fluke Calibration, American Fork- Primary Temperature Lab 799 E. Utah Valley Drive American Fork, UT 84003-9775 Mr. Thomas J. Wiandt Phone: 801-763-1700 Fax: 801-763-1010 E-mail: tom.wiandt@flukecal.com URL: http://us.flukecal.com/support/accreditations</p>	<p>Parameter(s) of Accreditation Electromagnetics - DC/Low Frequency Thermodynamic</p> <p>This laboratory is compliant to ANSI/NCSL Z540-1-1994; Part 1. (NVLAP Code: 20/A01)</p>
--	--

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
ELECTROMAGNETICS – DC/LOW FREQUENCY			
NVLAP Code: 20/E05 DC RESISTANCE AND CURRENT			
DC Resistance	1 Ω to 10 Ω >10 Ω to 100 Ω >100 Ω to 1 k Ω >1 k Ω to 10 k Ω >10 k Ω to 100 k Ω >100 k Ω to 500 k Ω >500 k Ω to 1 M Ω	0.30 $\mu\Omega/\Omega$ 0.40 $\mu\Omega/\Omega$ 0.55 $\mu\Omega/\Omega$ 0.65 $\mu\Omega/\Omega$ 1.5 $\mu\Omega/\Omega$ 2.0 $\mu\Omega/\Omega$ 3.0 $\mu\Omega/\Omega$	
Digital Thermometry Indicators			Readout devices that actually measure resistance Ratio is dimensionless
Ratio Function	0.25 to 4.0	0.20	
Resistance Function	1 Ω 10 Ω 25 Ω 100 Ω 10 k Ω 40 k Ω 100 k Ω 300 k Ω 500 k Ω	5.0 $\mu\Omega/\Omega$ 1.0 $\mu\Omega/\Omega$ 1.0 $\mu\Omega/\Omega$ 0.80 $\mu\Omega/\Omega$ 1.2 $\mu\Omega/\Omega$ 2.0 $\mu\Omega/\Omega$ 5.0 $\mu\Omega/\Omega$ 5.0 $\mu\Omega/\Omega$ 5.0 $\mu\Omega/\Omega$	

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
	0 Ω to 1 Ω >1 Ω to 25 Ω >25 Ω to 400 Ω >400 Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω	10 $\mu\Omega$ 10 $\mu\Omega/\Omega$ 4.0 $\mu\Omega/\Omega$ 8.0 $\mu\Omega/\Omega$ 8.0 $\mu\Omega/\Omega$ 25 $\mu\Omega/\Omega$	
DC Current – Source/Measure	4.0 mA to 20.0 mA	40 $\mu\text{A}/\text{A} + 0.4 \mu\text{A}$	
NVLAP Code: 20/E06 DC VOLTAGE Digital Thermometry Indicators Voltage Function	0 mV to 50 mV >50 mV to 100 mV	0.45 μV 0.75 μV	Readout devices that actually measure voltage.
Digital Thermometers Reference Junction Compensation Circuits	0.0 $^{\circ}\text{C}$ to 25.0 $^{\circ}\text{C}$	10 mK	Internal Reference Junction Compensation
THERMODYNAMIC			
NVLAP Code: 20/T02 HUMIDITY Humidity – Calibration of Digital Thermo-Hygrometers and Chilled Mirror Hygrometers RH Function	10 % RH to 50 % RH >50 % RH to 70 % RH >70 % RH to 90 % RH	0.30 % RH 0.35 % RH 0.40 % RH	Valid from 15 $^{\circ}\text{C}$ to 35 $^{\circ}\text{C}$
Temperature Function, in Air	0 $^{\circ}\text{C}$ to 70 $^{\circ}\text{C}$	0.025 $^{\circ}\text{C}$	
Temperature Function, in Stirred Liquid	0 $^{\circ}\text{C}$ to 70 $^{\circ}\text{C}$	0.010 $^{\circ}\text{C}$	

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
NVLAP Code: 20/T06 RADIATION THERMOMETRY Direct Calibration of Infrared Radiation Thermometers Using Blackbody Cavities	-15 °C	0.13 °C	Calibration is performed over the spectral band of 8 μm to 14 μm.
	0 °C	0.12 °C	
35 °C	0.12 °C		
50 °C	0.12 °C		
100 °C	0.12 °C		
120 °C	0.12 °C		
200 °C	0.12 °C		
250 °C	0.15 °C		
350 °C	0.23 °C		
500 °C	0.37 °C		
Direct Calibration of Infrared Calibrators Using Infrared Radiation Thermometers	-15 °C	0.20 °C	
	0 °C	0.16 °C	
	50 °C	0.16 °C	
	100 °C	0.20 °C	
	120 °C	0.21 °C	
	200 °C	0.27 °C	
	250 °C	0.32 °C	
	350 °C	0.45 °C	
500 °C	0.79 °C		
NVLAP Code: 20/T07 RESISTANCE THERMOMETRY Fixed Point Schedule 1 Direct Comparison	-197 °C	0.60 mK	

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson
For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
Fixed Point	-38.8344 °C	0.35 mK	TPHg
	0.010 °C	0.15 mK	TPW
	29.7646 °C	0.35 mK	MPGa
	156.599 °C	0.65 mK	FPIn
	231.928 °C	0.85 mK	FPSn
	419.527 °C	1.1 mK	FPZn
	660.323 °C	1.7 mK	FPAI
	961.78 °C	4.7 mK	FPAg
Fixed Point Schedule 2 Routine Measurement Capability	-197 °C	0.75 mK	LN ₂
	-38.8344 °C	0.75 mK	TPHg
	0.010 °C	0.50 mK	TPW
	29.7646 °C	0.75 mK	MPGa
	156.599 °C	1.5 mK	FPIn
	231.928 °C	1.5 mK	FPSn
	419.527 °C	1.8 mK	FPZn
	660.323 °C	2.5 mK	FPAI
961.78 °C	6.0 mK	FPAg	
Fixed Point Schedule 3 Single Power Calibration – any SPRTs	-197 °C	2.0 mK	LN ₂
	-38.8344 °C	2.0 mK	TPHg
	0.010 °C	2.0 mK	TPW
	29.7646 °C	2.0 mK	MPGa
	156.599 °C	3.0 mK	FPIn
	231.928 °C	3.0 mK	FPSn
	419.527 °C	4.0 mK	FPZn
	660.323 °C	7.0 mK	FPAI
961.78 °C	10 mK	FPAg	

2011-10-01 through 2012-09-30
Effective dates

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty (k=2) ^{Note 3}	Remarks
Fixed Point Schedule 4 – Single Power Calibration in Mini Fixed Points Comparison- High Quality PRTs	-197 °C	6.0 mK	LN ₂
	-100 °C	6.0 mK	
	-38.8344 °C	6.0 mK	TPHg
	0.010 °C	4.0 mK	TPW
	156.599 °C	6.0 mK	FPIn
	231.928 °C	7.0 mK	FPSn
	419.527 °C	9.0 mK	FPZn
	660.323 °C	14 mK	FPAI
	961.78 °C	30 mK	FPAg
Comparison Schedule 1 – Single Power Calibration by Comparison High Quality PRTs	-200 °C	10 mK	
	-100 °C to -50 °C	10 mK	
	-50 °C to 0 °C	8.0 mK	
	0.010 °C	6.0 mK	
	0 °C to 200 °C	9.0 mK	
	200 °C to 300 °C	12 mK	
	300 °C to 400 °C	14 mK	
	400 °C to 500 °C	16 mK	
Comparison Schedule 2 – Single Power Calibration by Comparison Any Quality PRTs	-200 °C	25 mK	
	-100 °C to -50 °C	25 mK	
	-50 °C to 0 °C	25 mK	
	0 °C to 100 °C	25 mK	
	100 °C to 300 °C	30 mK	
	300 °C to 420 °C	35 mK	

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson
For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
Comparison Schedule 3 – Dry Block Calibration of PRTs	420 °C to 500 °C	45 mK	Comparison in LN ₂ Comparison in Dry Block
	660.323 °C	50 mK	
	-200 °C	25 mK	
	-38 °C to 0 °C	25 mK	
	0 °C to 420 °C	45 mK	
Thermistors Precision Thermistors	420 °C to 660 °C	50 mK	
	-20 °C to 100 °C	1.5 mK	
	100 °C to 150 °C	3.0 mK	
Thermistors	-50 °C to -20 °C	5.0 mK	
	-20 °C to 120 °C	4.0 mK	
	120 °C to 150 °C	6.0 mK	
Certification of Thermometric Fixed Point Cells Direct Comparison to Reference Cells	TPHg	0.20 mK	
	TPW	0.07 mK	
	MPGa	0.07 mK	
	FPIIn	0.40 mK	
	FPSn	0.50 mK	
	FPZn	0.60 mK	
	FPAI	1.0 mK	
	FPAg	1.9 mK	
Direct Comparison to Working Cells	TPHg	0.25 mK	
	TPW	0.10 mK	
	MPGa	0.10 mK	
	FPIIn	0.60 mK	
	FPSn	0.75 mK	

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson
For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty (k=2) ^{Note 3}	Remarks
Dry Block Calibrators	FPZn	0.75 mK	Direct Comparison to PRT
	FPAI	1.2 mK	
	FPAg	2.5 mK	
	-45.0 °C to 155.0 °C	0.020 °C	
	155 °C to 225 °C	0.030 °C	
	225 °C to 425 °C	0.040 °C	
	425 °C to 660 °C	0.050 °C	
NVLAP Code: 20/T08 TEMPERATURE INDICATORS High Quality Thermocouples		Calibration of used thermocouples may result in larger uncertainties due to increased inhomogeneity.	
Type S and Type R	156.599 °C	0.10 °C	FPIn
	231.928 °C	0.10 °C	FPSn
	419.527 °C	0.13 °C	FPZn
	660.323 °C	0.18 °C	FPAI
	961.78 °C	0.25 °C	FPAg
Au/PT	0.010 °C	10 mK	TPW
	156.599 °C	20 mK	FPIn
	231.928 °C	20 mK	FPSn
	419.527 °C	20 mK	FPZn
	660.323 °C	25 mK	FPAI
	961.78 °C	35 mK	FPAg
	1000.00 °C	40 mK	Extrapolated
Digital Thermometer with PRT System		Probe uncertainty is not included in stated uncertainty.	Comparison or Mini Fixed Points
	-200 °C	10 mK	
	-100 °C to -50 °C	10 mK	
	-50 °C to 0 °C	8.0 mK	

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson
For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
	0.010 °C	5.0 mK	
	0 °C to 200 °C	8.0 mK	
	200 °C to 300 °C	9.0 mK	
	300 °C to 400 °C	10 mK	
	400 °C to 500 °C	11 mK	
	660.323 °C	15 mK	FPAI
END			

2011-10-01 through 2012-09-30
Effective dates

David E. Alderson

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200348-0

Scope Revised: 2012-02-02

Notes

Note 1: A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

Note 2: Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

Note 3: The uncertainty associated with a measurement in a CMC is an expanded uncertainty using a coverage factor, $k = 2$, with a level of confidence of approximately 95 %. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

Note 3a: The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

Note 3b: As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

Note 3c: As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.1.h. of NIST Handbook 150, Procedures and General Requirements.


Note 4: Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

Note 5: Calibration values listed with percent (%) are percent of reading or generated value unless otherwise noted.

Note 6: NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

Note 7: See NIST Handbook 150 for further explanation of these notes.

2011-10-01 through 2012-09-30
Effective dates


For the National Institute of Standards and Technology