



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

MDS Nordion QC Laboratory

447 March Road

Kanata Ontario K2K 1X8

CANADA

Mr. Kevin O'Hara

Phone: 613-592-3400 x2524 Fax: 613-591-7422

E-mail: kevin.ohara@mdsinc.com

URL: <http://www.mds.nordion.com>

CALIBRATION LABORATORIES

NVLAP LAB CODE 200370-0

IONIZING RADIATION DOSIMETRY

NVLAP Code: 20/102

High-Dose Dosimetry

Evaluation of Transfer-standard and Reference-standard Dosimeters for Production Irradiators

<i>Range</i>	<i>Best Uncertainty (\pm)^{note 1}</i>	<i>Remarks</i>
0.5 kGy to 50 kGy	3.6%	Results reported as absorbed dose.

Irradiation of Dosimeters for Known Absorbed-dose Levels

<i>Range</i>	<i>Best Uncertainty (\pm)^{note 1}</i>	<i>Remarks</i>
10 Gy to 250 kGy	2.3 %	Results reported as absorbed-dose values or may be analyzed for dosimeter response and reported as a calibration curve.

Calibration of Routine Dosimeters using Reference-standard or Transfer-standard Dosimeters

<i>Range</i>	<i>Best Uncertainty (\pm)^{note 1}</i>	<i>Remarks</i>
0.5 kGy to 50 kGy	4.0 %	Results reported as absorbed-dose values or may be analyzed for dosimeter response and reported as a calibration curve.

2009-04-01 through 2010-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200370-0

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.

2009-04-01 through 2010-03-31

Effective dates

A handwritten signature in cursive script that reads 'Sally S. Bruce'.

For the National Institute of Standards and Technology