



**National Voluntary  
Laboratory Accreditation Program**



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005**

**GeoCorp Industrial Controls, Inc.**  
 9010 River Road  
 Huron, OH 44839  
 Mr. Niel Fleetwood  
 Phone: 419-433-1101 Fax: 419-433-1102  
 E-mail: nfleetwood@geocorpinc.com  
 URL: http://www.geocorpinc.com

**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200496-0**

**THERMODYNAMICS**

*NVLAP Code:* 20/T08  
 Thermocouples

<i>Range in °C</i>	<i>Best Uncertainty (±) in °C<sup>note 1</sup></i>	<i>Remarks</i>
450 to 1093	1.42	Type B
> 1093 to 1454	2.18	Type B
0 to 1000	1.23	Type E
0 to 760	1.22	Type J
0 to 1093	1.23	Type K
> 1093 to 1327	2.06	Type K
0 to 1093	1.23	Type N
> 1093 to 1300	2.06	Type N
0 to 1093	1.42	Type R
> 1093 to 1454	2.18	Type R
0 to 1093	1.42	Type S
> 1093 to 1454	2.18	Type S
0 to 400	1.22	Type T

1. Represents an expanded uncertainty using a coverage factor, k = 2, at an approximate level of confidence of 95 %.

2009-10-01 through 2010-09-30

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*