



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Asanuma Giken Co., Ltd.

4079-1 Koto-Cho, Nishi-Ku Hamamatsu-shi
Shizuoka-ken 431-1103
JAPAN

Mr. Susumu Asanuma

Phone: 81-53-486-1110 Fax: 81-53-486-1109

E-mail: info@asanuma-g.jp

URL: http://www.asanuma-g.jp/

CALIBRATION LABORATORIES

NVLAP LAB CODE 200536-0

DIMENSIONAL

NVLAP Code: 20/D16

Coordinate Measuring Machines

Range

0 mm to 150 mm

0 in to 5.91 in

Best Uncertainty (\pm)^{note 1}

3.0 μ m

118.1 μ in

Remarks

Artifact Calibration

Artifact Calibration

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.

2009-04-01 through 2010-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology