



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Bios International Corporation

10 Park Place
Butler, NJ 07405-1371
Mr. Harvey Padden
Phone: 973-492-8400 x13 Fax: 973-492-8270
E-mail: padh@biosint.com
URL: www.biosint.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200661-0

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

DC/LOW FREQUENCY ELECTROMAGNETICS

NVLAP Code: 20/E05
DC Current

<i>Range</i>	<i>Best Uncertainty</i> (\pm) ^{note 1}	<i>Remarks</i>
0.1 to 20 mA	0.015 %	

NVLAP Code: 20/E06
DC Voltage

<i>Range</i>	<i>Best Uncertainty</i> (\pm) ^{note 1}	<i>Remarks</i>
0.1 to 28 V	0.015 %	

TIME AND FREQUENCY

NVLAP Code: 20/F01
Frequency Dissemination ^{note 2}

<i>Range</i>	<i>Best Uncertainty</i> (\pm) ^{note 1}	<i>Remarks</i>
0.1 Hz to 10 MHz	0.000025 %	Frequency Period 200 ns to 10 sec

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200661-0

MECHANICAL

NVLAP Code: 20/M05

Flow Rate

Range in sccm	Best Uncertainty (\pm)^{note 1}	Remarks
1.0 to 2.5	0.163 %	
2.5 to 5.0	0.127 %	
5.0 to 50 000	0.071 %	

THERMODYNAMIC

NVLAP Code: 20/T05

Pressure^{note 2}

Range	Best Uncertainty (\pm)^{note 1}	Remarks
0 kPa to 1 kPa	0.15 kPa	
87 kPa to 173 kPa	3.47 Pa	

NVLAP Code: 20/T07

Resistance Thermometry^{note 2}

Range in °C	Best Uncertainty (\pm)^{note 1}	Remarks
-20 to -5	0.05 °C	
-5 to 70	0.03 °C	
70 to 130	0.03 °C	

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. Calibration service provided in support of Bios International Corporation manufactured flow standards only.

2009-10-01 through 2010-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology