



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

GE Sensing

1100 Technology Park Drive
Billerica, MA 01821
Mr. Stephen Rehn
Phone: 978-437-1000 Fax: 502-479-6722
E-mail: stephen.rehn@ge.com
URL: www.gesensing.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200803-0

THERMODYNAMICS

NVLAP Code: 20/T02

Humidity

Generated in RH chamber @ 25° C

Range in % RH

25 to 85

Best Uncertainty (±) in % RH^{note 1}

1.1

NVLAP Code: 20/T05

Pressure

Pressure by comparison

Range in psia

0 to 150

Best Uncertainty (±) in psi^{note 1}

0.02

NVLAP Code: 20/T07

Resistance Thermometry

RTD Temperature measurement by comparison to SPRT

Temperature in °C

-196

0

100

390

Best Uncertainty (±) in mK^{note 1}

10

5.1

8

12

2009-04-01 through 2010-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200803-0

Temperature by comparison to RTD Probe

<i>Temperature in °C</i>	<i>Best Uncertainty (±) in mK ^{note 1}</i>
-60	28
0	28
100	28
400	28

Temperature Measurement by comparison to RTD Probe

<i>Temperature Range in °C</i>	<i>Best Uncertainty (±) in mK ^{note 1}</i>
-60 to 400	28

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.

2009-04-01 through 2010-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology