

**United States Department of Commerce
National Institute of Standards and Technology**

~ **Certificate of Measurement Traceability** ~
for

Los Angeles County

a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)

2009

The State standards are traceable to NIST and the laboratory has demonstrated* measurement proficiency through training and interlaboratory comparisons compliant to NIST HB 143:2007 (ISO/IEC 17025:2005).



*Carol T. Hockert,
for the National Institute of Standards and Technology*

Chief, Weights and Measures Division

Title

January 1, 2009 - December 31, 2009

Effective Date

***Demonstrated
Areas of Proficiency**

Mass

Echelon III

20 kg - 1 mg

5000 lb - 0.001 lb

Volume

Gravimetric, I

1 gal - 8 fl oz

Volume transfer, II

400 gal - 1 gal

