

United States Department of Commerce
National Institute of Standards and Technology

~ Certificate of Measurement Traceability ~
for

Tennessee

a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)

2008

The State standards are traceable to NIST and the laboratory
has *conditionally* demonstrated* measurement proficiency
through training and interlaboratory comparisons.



***Demonstrated
Areas of Proficiency**

Mass

Echelon III

100 g - 1 mg

1000 lb - 0.001 lb

8 oz - 0.03125 oz

Volume

Volume transfer, II

100 gal - 1 gal

Conditional Status:

Facility has inadequate environmental controls (NIST HB 143 5.3, 6.2). Conditional Recognition is issued only so the laboratory can provide services to meet weights and measures regulations.

Handwritten signature of Carol T. Hockert in black ink.

Carol T. Hockert,
for the National Institute of Standards and Technology

Chief, Weights and Measures Division

Title

January 1, 2008

Effective Date