



CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

<p>Oklahoma Bureau of Standards 2800 North Lincoln Boulevard Oklahoma City, OK 73105-4298 Mr. Robert Whitney Phone: 405-522-0316 Fax: 405-522-5457 E-mail: robert.whitney@ag.ok.gov URL: http://www.state.ok.us/lab-bos.htm</p>	<p>Parameter(s) of Accreditation Mechanical</p> <p>This laboratory is compliant to ANSI/NC SL Z540-1-1994; Part 1. (NVLAP Code: 20/A01)</p>
---	--

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) <small>Note 3</small>	Remarks
MECHANICAL			
<p>NVLAP Code: 20/M08 MASS Metric</p>	<p>20 kg 10 kg 5 kg 3 kg 2 kg 1 kg 500 g 300 g 200 g 100 g 50 g 30 g 20 g 10 g 5 g 3 g 2 g 1 g 500 mg 300 mg 200 mg 100 mg 50 mg</p>	<p>7.1 mg 0.79 mg 0.53 mg 0.46 mg 0.44 mg 52 µg 27 µg 18 µg 13 µg 11 µg 6.3 µg 4.3 µg 3.4 µg 3.3 µg 1.7 µg 1.2 µg 0.88 µg 0.83 µg 0.57 µg 0.45 µg 0.42 µg 0.48 µg 0.26 µg</p>	<p>Echelon I</p>

2012-01-01 through 2012-12-31
 Effective dates

David F. Alderson
 For the National Institute of Standards and Technology