





**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200605-0**

*NVLAP Code:* 20/D05  
Length & Diameter - Indicators

<i>Range in inches</i>	<i>Best Uncertainty (±) in µin <sup>note 1</sup></i>	<i>Remarks</i>
up to 0.100	21	M&TE
0.100 to 0.250	82	M&TE
0.250 to 2	82	M&TE

*NVLAP Code:* 20/D05  
Length - Air Amplifiers

<i>Range in inches</i>	<i>Best Uncertainty (±) in µin <sup>note 1</sup></i>	<i>Remarks</i>
0.0003 to 0.003	12	M&TE Dimensionair®

*NVLAP Code:* 20/D05  
Length

<i>Range in inches</i>	<i>Best Uncertainty (±) in µin <sup>note 1</sup></i>	<i>Remarks</i>
0.0003 to 0.003	13	M&TE All Mahr Federal Inc. AMR Kits

*NVLAP Code:* 20/D05  
Length

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>	<i>Remarks</i>
< 400 arc seconds	0.40 arc seconds	M&TE Electronic Levels System
0 in to 1 in	58 µin	M&TE 400 B3 & B4 Calibrators

*NVLAP Code:* 20/D05  
Length & Diameter - Outside Micrometers

<i>Range in inches</i>	<i>Best Uncertainty (±) in µin <sup>note 1</sup></i>	<i>Remarks</i>
0 to 1	58	M&TE
1 to 2	58	M&TE

2011-04-01 through 2012-03-31

*Sally S. Bruce*

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200605-0**

2 to 3	58	M&TE
3 to 4	58	M&TE
4 to 5	58	M&TE
5 to 6	58	M&TE

**NVLAP Code:** 20/D05  
Length – Field Service Calibration

<b>Range</b>	<b>Best Uncertainty (<math>\pm</math>) <sup>note 1</sup></b>	<b>Remarks</b>
Universal Length Measuring Machines 0.5 in to 12 in	5.0 $\mu$ in (0.127 $\mu$ m)	Gage Blocks
Universal Height Measuring Machines 5 mm to 700 mm	80 $\mu$ in (2.0 $\mu$ m)	Calibrated Step Gage
Universal Calibrators 0.5 in	9.0 $\mu$ in (229 $\mu$ m)	Gage Blocks
Comparators 0.002 in	3.1 $\mu$ in (0.079 $\mu$ m)	Gage Blocks

**NVLAP Code:** 20/D09  
Roundness

<b>Range</b>	<b>Best Uncertainty (<math>\pm</math>) in <math>\mu</math>in <sup>note 1</sup></b>	<b>Remarks</b>
0.124 in to 2 in Dia. with a roundness <100 $\mu$ in	1 $\mu$ in	
0.124 in to 14.5 in Dia. with a roundness $\leq$ 0.004 in	3.5 $\mu$ in (0.089 $\mu$ m)	
0.124 in to 14.5 in Dia. with a roundness > 0.004 to 0.40 in	25 $\mu$ in (0.64 $\mu$ m)	

2011-04-01 through 2012-03-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200605-0**

**NVLAP Code:** 20/D11  
Spherical Diameter; Plug

<b>Range in inches</b>	<b>Best Uncertainty (<math>\pm</math>) in <math>\mu\text{in}</math> <sup>note 1</sup></b>	<b>Remarks</b>
up to 1	6	
1 to 2	7	
2 to 4	10	
4 to 10	(10 + 1L)	

**NVLAP Code:** 20/D11  
Ring Gages

<b>Range in inches</b>	<b>Best Uncertainty (<math>\pm</math>) in <math>\mu\text{in}</math> <sup>note 1</sup></b>	<b>Remarks</b>
0.125 to 5.0	7	Mahr 828 CIM
up to 1	6	
1 to 2	7	
2 to 4	10	
4 to 14	(10 + 1L)	

**NVLAP Code:** 20/D11  
Air Rings

<b>Range in inches</b>	<b>Best Uncertainty (<math>\pm</math>) in <math>\mu\text{in}</math> <sup>note 1</sup></b>	<b>Remarks</b>
< 2	18	M&TE
2 to 4	25	M&TE

**NVLAP Code:** 20/D11  
Air Plugs

<b>Range in inches</b>	<b>Best Uncertainty (<math>\pm</math>) in <math>\mu\text{in}</math> <sup>note 1</sup></b>	<b>Remarks</b>
< 1	12	M&TE
$\geq$ 1 to 2	26	M&TE
> 2 to 3	28	M&TE
> 3 to 4	32	M&TE
> 4 to 5	33	M&TE

2011-04-01 through 2012-03-31

*Sally S. Bruce*

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200605-0**

*NVLAP Code:* 20/D12  
Surface Texture

<i>Range</i>	<i>Best Uncertainty (±) in μin<sup>note 1</sup></i>	<i>Remarks</i>
20 μin R <sub>a</sub> to 300 μin R <sub>a</sub>	2	

*NVLAP Code:* 20/D12  
Surface Finish / Contour Measuring Machines – Field Service Calibration

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>	<i>Remarks</i>
R <sub>a</sub> 100 μin to 150 μin	1.18 μin (0.03 μm)	Surface Finish Standard
W <sub>t</sub> <60 μin/in.	3.15 μin (0.08 μm)	Straight Edge
Displacement 180 μin to 240 μin	3.0 μin (0.076 μm)	Step Height Standard
Probe Calibration Steps 1 mm to 70 mm	15.8 μin (0.40 μm)	Gage Blocks
Gage Pin Radius 2 mm to 4 mm	5.12 μin (0.13 μm)	Calibrated Gage Pin
Sphere Radius >4mm to 25 mm	5.12 μin (0.13 μm)	Calibrated Sphere (2 ball master)

2011-04-01 through 2012-03-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0

*NVLAP Code:* 20/D15

Precision Geometry / Form Measuring Machines

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
<b>Concentricity</b>		
$\leq 14.5$ in Dia. and $\leq 13.75$ in. Hgt. with a concentricity of $\leq 0.004$ in	12 $\mu$ in (0.3 $\mu$ m)	
$\leq 14.5$ in Dia. and $\leq 13.75$ in. Hgt with a concentricity of $> 0.004$ in to 0.040 in to 0.0040 in	27 $\mu$ in (0.69 $\mu$ m)	
<b>Cylindricity</b>		
$\leq 1.0$ in Hgt. and $\leq 14.5$ in Dia. with a cylindricity of $\leq 0.0001$ in	5 $\mu$ in (0.13 $\mu$ m)	
$\leq 4.0$ in Hgt. and $\leq 14.5$ in Dia. with a cylindricity of $\leq 0.004$ in	15 $\mu$ in (0.38 $\mu$ m)	
$> 4.0$ in to 13.75 in Hgt. and $\leq 14.5$ in Dia. with a cylindricity of $\leq 0.004$ in	25 $\mu$ in (0.64 $\mu$ m)	
$\leq 4.0$ in Hgt. and $\leq 14.5$ in Dia. with a cylindricity of $> 0.004$ in to 0.040 in to 0.040 in	29 $\mu$ in (0.74 $\mu$ m)	
$\geq 4.0$ in to 13.75 in Hgt. and $\leq 14.5$ in Dia. with a cylindricity of $> 0.004$ in to 0.040 in A	35 $\mu$ in (0.89 $\mu$ m)	
<b>Flatness</b>		
$\leq 14.5$ in Dia. $\leq 13.75$ in Hgt. with a flatness of $\leq 0.004$ in	3 $\mu$ in (0.08 $\mu$ m)	
$\leq 14.5$ in Dia. $\leq 13.75$ in Hgt. with a flatness of $> 0.004$ in to 0.040 in	25 $\mu$ in (0.64 $\mu$ m)	
<b>Parallelism</b>		
$\leq 14.5$ in Dia. $\leq 13.75$ in Hgt. with a parallelism of $\leq 0.004$ in	4 $\mu$ in (0.10 $\mu$ m)	
$\leq 14.5$ in Dia. $\leq 13.75$ in Hgt. with a parallelism of $> 0.004$ in to 0.040 in	25 $\mu$ in (0.64 $\mu$ m)	

2011-04-01 through 2012-03-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0

Perpendicularity  
 ≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a  
 Perpendicularity of ≤ 0.004 in 5 μin  
 (0.13 μm)  
 ≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a  
 perpendicularity of > 0.004 in to 0.040 in 25 μin  
 (0.64 μm)

Runout  
 ≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a  
 runout of ≤ 0.004 in 4 μin  
 (0.1 μm)  
 ≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a  
 runout of > 0.004 in to 0.040 in 25 μin  
 (0.64 μm)

Total Runout  
 ≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a total  
 runout of ≤ 0.004 in 330 μin  
 (8.4 μm)

**NVLAP Code:** 20/D15  
 Geometry / Form Measuring Machines – Field Service Calibration

<b>Range</b>	<b>Best Uncertainty (±) <sup>note 1</sup></b>	<b>Remarks</b>
Radial Departure <50 μin	1.85 μin (0.047 μm)	Precision Sphere
Axial Deviation <50 μin	1.2 μin (0.03 μm)	Optical Flat
Coning Error <10 μin / in	1.17 μin (0.03 μm)	Precision Sphere
Probe Calibration <0.040 in	26.4 μin (0.67 μm)	Gage Blocks

2011-04-01 through 2012-03-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200605-0**

Straightness	<2 $\mu\text{m}$ / 100 mm	5.9 $\mu\text{in}$ (0.15 $\mu\text{m}$ )	Straight Edge
Z Axis Parallelism	<10 $\mu\text{m}$ / m	87.4 $\mu\text{in}$ (2.22 $\mu\text{m}$ )	Cylindrical Square
X Axis Perpendicular	<10 $\mu\text{m}$ /m	281 $\mu\text{in}$ (7.74 $\mu\text{m}$ )	Straight Edge

- 
1. Represents an expanded uncertainty using a coverage factor,  $k = 2$ , at an approximate level of confidence of 95 %.
  2. L in inches
  3. L in mm
  4. Add 1.5  $\mu\text{in}$ /38.1 nm for chrome carbide, 2.3  $\mu\text{in}$ /58.4 nm for tungsten carbide to the uncertainty listed. The uncertainty listed is for steel gage blocks.

2011-04-01 through 2012-03-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*