



**National Voluntary  
Laboratory Accreditation Program**



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005**

**Rainin Instrument LLC Oakland, CA Pipette Service**

7500 Edgewater Drive  
Oakland, CA 94621-0060  
Mr. Bruce Loughridge  
Phone: 510-564-1608 Fax: 510-564-1757  
E-mail: bruce.loughridge@rainin.com  
URL: <http://rainin.com>

**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200638-0**

*NVLAP Code:* 20/A01                      ANSI/NCSL Z540-1-1994; Part 1                      Compliant

**MECHANICAL**

*NVLAP Code:* 20/M12  
Volume Pipettes

Gravimetric Method

<i>Pipette Nominal Volume in <math>\mu</math></i>	<i>% of Nominal Volume</i>	<i>Best Uncertainty (<math>\pm</math>) in <math>\mu</math>L <sup>note 1</sup></i>
2	10	0.014
2	50	0.014
2	100	0.016
10	10	0.013
10	50	0.016
10	100	0.022
20	10	0.038
20	50	0.044
20	100	0.044
100	10	0.18
100	50	0.21
100	100	0.26
200	10	0.17
200	50	0.25

2011-04-01 through 2012-03-31

*Sally S. Bruce*

*Effective dates*

*For the National Institute of Standards and Technology*



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200638-0**

200	100	0.38
1000	10	0.94
1000	50	1.6
1000	100	2.2
2000	10	1.4
2000	50	2.8
2000	100	3.5
5000	10	3.2
5000	50	6.5
5000	100	7.1
10 000	10	5.9
10 000	50	10.7
10 000	100	13.2
20 000	10	12.3
20 000	50	17.5
20 000	100	22.4

1. Represents an expanded uncertainty using a coverage factor,  $k = 2$ , at an approximate level of confidence of 95%.

2011-04-01 through 2012-03-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*