



National Voluntary Laboratory Accreditation Program



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

U.S. Army Radiation Standards Laboratory / Radiation Standards Lab.

Army Primary Standards Laboratory Directorate

Attn: AMSAM-TMD-SR

Redstone Arsenal, AL 35898-5000

Mr. Patrick Kuykendall

Phone: 256-876-5593/3576 Fax: 256-955-6413

E-mail: Patrick.kuykendall@us.army.mil

CALIBRATION LABORATORIES

NVLAP LAB CODE 200715-0

NVLAP Code: 20/A01

ANSI/NCSL Z540-1-1994; Part 1

Compliant

IONIZING RADIATION

NVLAP Code: 20/I01

Dosimetry of X-Rays, Gamma Rays, and Electrons

Calibration of Survey Instruments

Calibration Category	Radiation Type	Emission Rate Range	Reference Field Uncertainty (\pm) in % notes 1,2
Alpha	^{238}Pu	2×10^6 counts / min	2.5
Calibration Category	Radiation Type	Exposure Rate Range note 3	Reference Field Uncertainty (\pm) in % notes 1,2
Gamma	^{137}Cs	0.05 mR/hr to 5000 R/hr	3.0
X-Ray	M30	20 to 200 R/hr	3.0
	M60	1.5 to 300 R/hr	3.0
	M100	1.5 to 300 R/hr	3.0
	M150	2.5 to 400 R/hr	3.0
	H150	0.3 to 15 R/hr	3.0
	S75	50 to 500 R/hr	3.0
	HK10	15 to 100 R/hr	3.0
	HK20	25 to 250 R/hr	3.0

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200715-0

HK30	20 to 200 R/hr	3.0
HK60	4.0 to 150 R/hr	3.0
HK100	4.0 to 150 R/hr	3.0
HK200	10 to 400 R/hr	3.0
HK250	15 to 500 R/hr	3.0
HK280	15 to 500 R/hr	3.0
HK300	15 to 500 R/hr	3.0
LK10	0.15 to 0.8 R/hr	3.0
LK20	0.4 to 2.5 R/hr	3.0
LK30	0.04 to 0.4 R/hr	3.0
LK35	0.075 to 1.0 R/hr	3.0
LK55	0.015 to 0.5 R/hr	3.0
LK70	0.04 to 0.4 R/hr	3.0
LK100	0.01 to 0.4 R/hr	3.0
LK125	0.0008 to 0.4 R/hr	3.0
LK170	0.0005 to 0.4 R/hr	3.0
LK210	0.005 to 0.25 R/hr	3.0
LK240	0.0015 to 0.3 R/hr	3.0
NS10	1.5 to 10 R/hr	3.0
NS15	1.0 to 8.0 R/hr	3.0
NS20	1.5 to 15 R/hr	3.0
NS25	1.5 to 15 R/hr	3.0
NS30	1.0 to 10 R/hr	3.0
NS40	0.1 to 4.0 R/hr	3.0
NS60	0.2 to 8.0 R/hr	3.0
NS80	0.5 to 5.0 R/hr	3.0
NS100	0.04 to 2.0 R/hr	3.0
NS120	0.2 to 2.0 R/hr	3.0
NS150	0.3 to 10 R/hr	3.0
NS200	0.2 to 5.0 R/hr	3.0
NS250	0.15 to 5.0 R/hr	3.0
NS300	0.25 to 5.0 R/hr	3.0
WS60	0.5 to 20 R/hr	3.0
WS80	1.0 to 40 R/hr	3.0

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200715-0

WS110	0.75 to 25 R/hr	3.0
WS150	1.5 to 50 R/hr	3.0
WS200	2.0 to 100 R/hr	3.0
WS250	2.5 to 100 R/hr	3.0
WS300	3.0 to 150 R/hr	3.0

Irradiation of Personnel Dosimeters

Calibration Category	Radiation Type	Nominal Range	Reference Field Uncertainty (\pm) in % ^{notes 1,2}
Gamma	^{137}Cs	>0.03	3.0
X-Ray	M30	>0.03	3.0
	M60	>0.03	3.0
	M100	>0.03	3.0
	M150	>0.03	3.0
	H150	>0.03	3.0
	S75	>0.03	3.0
	HK10	>0.03	3.0
	HK20	>0.03	3.0
	HK30	>0.03	3.0
	HK60	>0.03	3.0
	HK100	>0.03	3.0
	HK200	>0.03	3.0
	HK280	>0.03	3.0
	HK300	>0.03	3.0
	LK10	>0.03	3.0
	LK20	>0.03	3.0
	LK30	>0.03	3.0
	LK35	>0.03	3.0
	LK55	>0.03	3.0
	LK70	>0.03	3.0
	LK100	>0.03	3.0
	LK125	>0.03	3.0
	LK170	>0.03	3.0

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200715-0

LK210	>0.03	3.0
LK240	>0.03	3.0
NS10	>0.03	3.0
NS15	>0.03	3.0
NS20	>0.03	3.0
NS25	>0.03	3.0
NS30	>0.03	3.0
NS40	>0.03	3.0
NS60	>0.03	3.0
NS80	>0.03	3.0
NS100	>0.03	3.0
NS120	>0.03	3.0
NS150	>0.03	3.0
NS200	>0.03	3.0
NS250	>0.03	3.0
NS300	>0.03	3.0
WS60	>0.03	3.0
WS80	>0.03	3.0
WS110	>0.03	3.0
WS150	>0.03	3.0
WS200	>0.03	3.0
WS250	>0.03	3.0
WS300	>0.03	3.0

Irradiation of Thermoluminescent Dosimeters and Electronic Dosimeters

Calibration Category	Radiation Source Type	Absorbed Dose Rate	Reference Field Uncertainty (\pm) in % ^{notes 1,2}
Beta	⁹⁰ SrY	(0.6 to 25) μ Gy/sec	2.5
	⁸⁵ Kr	(0.6 to 25) μ Gy/sec	2.5
	¹⁴⁷ Pm	(0.6 to 25) μ Gy/sec	2.5

NVLAP Code: 20/I04

Radioactive Sources

Radionuclide Source Calibration

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200715-0

Radioisotope	Nominal Range in Bq ^{note 3}	Best Uncertainty (\pm) in % ^{notes 1,2}
^{238}Pu	to 10^6	2.5
^{239}Pu	to 10^6	2.5
^{241}Am	to 10^6	2.5

Gamma / X-Ray Source Calibration for Air Kerma Rate

Radioisotope / Beam Code	Range ^{note 3}	Best Uncertainty (\pm) in % ^{notes 1,2}
^{137}Cs ^{note 4}	0.05 mR/hr to 5000 R/hr	3.0
^{137}Cs ^{notes 4 and 5}	0.1 mR/hr to 600 R/hr	4.3
^{60}Co	0.05 mR/hr to 1000 R/hr	3.0

NIST Beam Codes

M30, M60, M100 M150, H150, 0.3 to 400 R/hr 3.0

S75

ISO Beam Codes

All HK, LK, NS, WS 0.5 mR/to 500 R/hr 3.0

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. Uncertainties are valid for nominal intensity range listed.
3. For calibration outside of nominal intensity range shown, uncertainties would be determined commensurate with parameters of the reference field calibration.
4. This capability includes off-site calibration service, as limited by influences of operating environment.
5. Enclosed calibration range.

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology