



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

| | |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>Boeing Test & Evaluation, Metrology Lab Operations MC 2R – 40, PO Box 3707 Seattle, WA 98124-2207 Mr. Derek Porter Phone: 206-544-4885 Fax: 206-544-4923 E-mail: derek.j.porter@boeing.com URL: http://www.boeing.com/commercial/techsvcs/boeingtech</p> | <p>Parameter(s) of Accreditation Dimensional Electromagnetics – DC/Low Frequency Time and Frequency Mechanical Electromagnetics – RF/Microwave Thermodynamics</p> |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|-------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------|
| DIMENSIONAL | | | |
| <p>NVLAP Code: 20/D01 Angular</p> | <p>>20 s to 1000 s 0 s to 20 s 0 ° to 360 ° 0 minutes -10 Minutes 0 minutes -10 Minutes</p> | <p>0.25 Arc Seconds 0.20 Arc Seconds 0.4 Arc Second 0.5 Arc Second or 2 % (whichever is greater) 0.2% + 0.5 Arc Second Per Meter of Separation</p> | <p>Autocollimator Autocollimator Moore Indexing Table Small Angle Generator Laser Angular Optics</p> |
| <p>NVLAP Code: 20/D11 Plain Ring Gages</p> | <p>0.04 to 13.0 in</p> | <p>(7.0 + 0.4L) μin (where L is in inches)</p> | <p>Plain Ring Gages</p> |
| <p>NVLAP Code: 20/D03 Gage Blocks</p> | <p>< 0.05 in 0.050 in to 4.0 in 5 in to 20 in</p> | <p>2.5 μin (2.2 + 0.3L) μin (where L is in inches) (4.3 + 0.35L) μin (where L is in inches)</p> | <p>Laser / Mechanical comparison to NIST calibrated masters Mechanical comparison to NIST calibrated masters</p> |
| <p>NVLAP Code: 20/D05 Length & Diameter, Step Gages</p> | <p>>0 feet to 25 Feet >0 inch to 5 inches</p> | <p>1.4 μinch / inch +/-0.0001"</p> | <p>Laser Interferometric Bench Coordinate measuring machine</p> |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|----------------------------------------------|------------------------------------------------------------|-------------------------------------------------|------------------------------------------------------------|
| NVLAP Code: 20/D09 Roundness | 0.05 in to 3 in >3.0 in to 8.5 in >8.5 in to 11.8 in | 3.0 μ in 4.0 μ in 6.0 μ in | Uncertainties increase as height increase above turntable. |
| NVLAP Code: 20/D16 Laser Trackers | 40 meter radius | (0.001" + 10 μ inch/inch) (Line of Site) | Scale Bar |

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | | | | | | | | | |
|----------------------------------------------------------------------------------|---------------|-----------------------------------------------------------|-------|--------|--------|-------|-------|--------|--------|--------|---------|
| | | ELECTROMAGNETICS – DC/LOW FREQUENCY | | | | | | | | | |
| | | <i>(in μA/A for indicated frequencies)</i> | | | | | | | | | |
| | Current Level | 10 Hz | 20 Hz | 100 Hz | 400 Hz | 1 kHz | 5 kHz | 10 kHz | 20 kHz | 50 kHz | 100 kHz |
| NVLAP Code: 20/E01 VOLTAGE/ CURRENT CONVERTERS (to 1 MHz) | 10 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 77 | 110 |
| | 20 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 77 | 110 |
| | 30 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 77 | 110 |
| | 50 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 77 | 110 |
| | 100 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 77 | 110 |
| | 200 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 77 | 110 |
| | 300 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 88 | 110 |
| AC/DC | 500 mA | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 55 | 110 | 220 |
| Difference for Low Frequency Detectors and Shunts | 1 A | 160 | 55 | 55 | 55 | 55 | 55 | 55 | 59 | 130 | 250 |
| | 2 A | 160 | 66 | 55 | 55 | 55 | 55 | 55 | 59 | 130 | 250 |
| | 3 A | 160 | 77 | 99 | 66 | 66 | 140 | 140 | 160 | 300 | 550 |
| | 5 A | 160 | 88 | 99 | 88 | 88 | 88 | 140 | 190 | 610 | 1000 |
| | 10 A | 160 | 110 | 120 | 110 | 110 | 110 | 88 | 110 | 160 | |
| | 20 A | 180 | 150 | 140 | 110 | 110 | 110 | 110 | 130 | 253 | |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|-------------------------------------------------------------|----------------------------------|-----------------------------------------|-------------------------------------------------|
| NVLAP Code: 20/E05 DC RESISTANCE and CURRENT | 1 $\mu\Omega$ to 1 m Ω | 25 $\mu\Omega/\Omega$ | Using MI 6010 Bridge w/ 6011, 6012 Extenders |
| | >1 m Ω to 10 m Ω | 0.60 $\mu\Omega/\Omega$ | Using MI 6010 Bridge w/ 6011 extender |
| | >10 m Ω to 100 m Ω | 0.44 $\mu\Omega/\Omega$ | |
| | >0.1 Ω to <1 Ω | 0.36 $\mu\Omega/\Omega$ | Using MI 6010B Bridge |
| | 1 Ω | 0.30 $\mu\Omega/\Omega$ | |
| | >1 Ω to 10 Ω | 0.32 $\mu\Omega/\Omega$ | |
| | >10 Ω to 100 Ω | 0.33 $\mu\Omega/\Omega$ | |
| | >100 Ω to 1 k Ω | 0.35 $\mu\Omega/\Omega$ | |
| | 1 k Ω to 10 k Ω | 0.37 $\mu\Omega/\Omega$ | Using MI 6001A Bridge |
| | 10 k Ω to 100 k Ω | 1.1 $\mu\Omega/\Omega$ | |
| | 100 k Ω to 1 M Ω | 1.5 $\mu\Omega/\Omega$ | |
| | 1 M Ω to 10 M Ω | 1.8 $\mu\Omega/\Omega$ | |
| | 10 M Ω to 100 M Ω | 2.1 $\mu\Omega/\Omega$ | |
| | 100 M Ω to 1 G Ω | 75 $\mu\Omega/\Omega$ | Using Guildline 6520 Teraohmmeter |
| | 1 G Ω to 10 G Ω | 0.013 % | |
| 10 G Ω to 100 G Ω | 0.015 % | | |
| 100 G Ω to 1 T Ω | 0.030 % | | |
| 10 T Ω to 10 T Ω | 0.035 % | | |
| NVLAP Code: 20/E06 DC VOLTAGE Voltage | 1.0 V to 1.018 V | 0.4 $\mu\text{V}/\text{V}$ | Using 3 Zener Intercomparison |
| | 10 V | 0.19 $\mu\text{V}/\text{V}$ | |
| | 1.0 V to 1.018 V | 0.02 $\mu\text{V}/\text{V}$ | Using Josephson Junction |
| 10 V | 0.02 $\mu\text{V}/\text{V}$ | | |
| High Voltage – Dividers, Meters, & Indicators | 100 V to 100 kV | 0.013 % | DC Mode |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | | Remark | | | | | | | | | | | | |
|-----------------------------------------|--------|----------------------------------------------------------------------------------------------------------------------------------------|-------|-------|--------|-------|--------|--------|--------|---------|---------|---------|---------|-------|
| NVLAP Code: 20/E09 | | Direct Comparison Method: AC Source is supplied to both customer unit under test and Fluke 792A Thermal Transfer Standard in parallel. | | | | | | | | | | | | |
| LF AC VOLTAGE | | Uncertainty ($k=2$) ^{Note 3} (at indicated frequency, uncertainty values expressed in $\mu\text{V/V}$) | | | | | | | | | | | | |
| Range | Input | 10 Hz | 20 Hz | 40 Hz | 100 Hz | 1 kHz | 10 kHz | 20 kHz | 50 kHz | 100 kHz | 300 kHz | 500 kHz | 800 kHz | 1 MHz |
| 22 mV | 2 mV | 420 | 420 | 420 | 410 | 410 | 410 | 410 | 440 | 540 | 630 | 750 | 880 | 880 |
| 22 mV | 6 mV | 330 | 330 | 290 | 270 | 270 | 270 | 270 | 360 | 420 | 520 | 590 | 730 | 770 |
| 22 mV | 10 mV | 96 | 80 | 80 | 80 | 80 | 80 | 80 | 140 | 200 | 250 | 330 | 423 | 467 |
| 22 mV | 20 mV | 96 | 80 | 80 | 80 | 80 | 80 | 80 | 140 | 200 | 250 | 330 | 423 | 467 |
| 220 mV | 20 mV | 94 | 77 | 77 | 75 | 75 | 75 | 75 | 100 | 170 | 250 | 340 | 424 | 468 |
| 220 mV | 60 mV | 56 | 51 | 45 | 41 | 41 | 35 | 35 | 53 | 95 | 170 | 250 | 381 | 381 |
| 220 mV | 100 mV | 51 | 45 | 39 | 33 | 21 | 21 | 27 | 27 | 50 | 94 | 130 | 291 | 291 |
| 220 mV | 200 mV | 33 | 28 | 17 | 16 | 16 | 16 | 17 | 26 | 44 | 94 | 130 | 274 | 287 |
| 700 mV | 200 mV | 34 | 28 | 18 | 16 | 16 | 16 | 17 | 26 | 44 | 94 | 130 | 273 | 286 |
| 700 mV | 600 mV | 32 | 24 | 23 | 10 | 9 | 9 | 10 | 10 | 14 | 32 | 38 | 62 | 74 |
| 2.2 V | 600 mV | 32 | 26 | 22 | 10 | 8 | 8 | 9 | 11 | 12 | 28 | 32 | 44 | 54 |
| 2.2 V | 1.0 V | 32 | 26 | 22 | 10 | 8 | 8 | 8 | 10 | 11 | 27 | 33 | 45 | 54 |
| 2.2 V | 2.0 V | 32 | 26 | 22 | 10 | 8 | 8 | 8 | 8 | 10 | 27 | 33 | 45 | 54 |
| 7 V | 2.0 V | 32 | 26 | 22 | 10 | 8 | 8 | 8 | 9 | 10 | 26 | 32 | 38 | 50 |
| 7 V | 6.0 V | 32 | 26 | 22 | 10 | 8 | 8 | 8 | 9 | 10 | 26 | 32 | 38 | 50 |
| 22 V | 6.0 V | 32 | 27 | 23 | 10 | 8 | 8 | 8 | 10 | 10 | 27 | 33 | 50 | 59 |
| 22 V | 10 V | 32 | 26 | 22 | 10 | 8 | 8 | 8 | 9 | 11 | 26 | 32 | 50 | 59 |
| 22 V | 20 V | 32 | 26 | 22 | 10 | 9 | 9 | 9 | 10 | 12 | 26 | 32 | 50 | 59 |
| 70 V | 20 V | 32 | 26 | 22 | 10 | 9 | 9 | 9 | 10 | 14 | 65 | 69 | | |
| 70 V | 60 V | 33 | 27 | 26 | 10 | 9 | 9 | 9 | 11 | 14 | 32 | 36 | | |
| 220 V | 60 V | 32 | 26 | 23 | 10 | 9 | 9 | 9 | 17 | 20 | 34 | | | |
| 220 V | 100 V | 32 | 26 | 23 | 10 | 9 | 9 | 10 | 15 | 17 | | | | |
| 220 V | 200 V | 45 | 27 | 24 | 12 | 11 | 11 | 11 | 17 | 22 | | | | |
| 1000 V | 200 V | 50 | 27 | 24 | 15 | 14 | 14 | 14 | 21 | 40 | | | | |
| 1000 V | 600 V | 62 | 27 | 27 | 21 | 15 | 15 | 15 | 21 | 44 | | | | |
| 1000 V | 1000 V | 62 | 27 | 27 | 24 | 22 | 22 | 23 | 32 | 56 | | | | |

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|-----------------------------------------------|----------------|-----------------------------------------|-----------------|
| High Voltage – Dividers, Meters, & Indicators | 100 V to 69 kV | 0.02 % | AC Mode (60 Hz) |

2012-01-01 through 2012-12-31
Effective dates

David T. Alderman
For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Uncertainty ($k=2$) ^{Note 3} | | | | | |
|----------------------------------------------------|-----------------------------------------------------------------------------------|--------------|---------------|--------------|---------------|---------------|
| | (at indicated frequency, uncertainty values expressed in $\mu\text{F}/\text{F}$) | | | | | |
| <i>NVLAP Code: 20/E10</i> LF Capacitance | <i>Range</i> | 50 Hz | 100 Hz | 1 kHz | 10 kHz | 20 kHz |
| Direct Comparison with Fused Silica Standards | 1 pF | 45 | 15 | 6 | 22 | 75 |
| | 10 pF | 12 | 6 | 2 | 5 | 15 |
| | 100 pF | 12 | 6 | 2 | 5 | 10 |
| | 200 pF | 12 | 6 | 2 | 5 | 10 |

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | | | |
|-----------------------------------------------------------------|-----------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------|-----------------------|-----------------------|
| | | (in $\mu\text{F}/\text{F}$ at selected frequencies) * | | | |
| Direct Measurement Using Capacitance Bridge | >0 pF to 1.65 μF | 50 Hz 16 | 1 kHz 6 | 10 kHz 13 | 20 kHz 16 |
| Using AH 2700 Bridge | (50 Hz to 20 kHz) | * These are best case and vary depending on test voltage, dissipation, and capacitance being measured. Floor value varies from (5 to 30) aF and will significantly increase stated values as measurement approaches low end of range | | | |
| | | Uncertainty ($k=2$) ^{Note 3} | | | |
| | | (in % at indicated frequency) | | | |
| NVLAP Code: 20/E11 LF INDUCTANCE Inductance | 100 μH | 100 Hz 0.15 | 1 kHz 0.15 | 10 kHz 0.15 | 20 kHz 0.15 |
| | 1 mH | 0.050 | 0.050 | 0.050 | 0.050 |
| | 10 mH | 0.050 | 0.050 | 0.050 | 0.25 |
| | 100 mH | 0.050 | 0.050 | 0.050 | |
| | 200 mH | 0.050 | 0.050 | 0.050 | |
| | 1 H | 0.050 | 0.080 | 0.16 | |
| | 2 H | 0.050 | 0.080 | | |
| | 5 H | 0.050 | 0.050 | | |
| | 10 H | 0.050 | 0.050 | | |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|------------------------------------------------------------------------------------------|----------------------------------------|-------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| TIME and FREQUENCY | | | |
| NVLAP Code: 20/F01 FREQUENCY DISSEMINATION Frequency Stability | 10 MHz | 1 part in 10^{11} 1 part in 10^{11} 4 parts in 10^{12} 1 part in 10^{12} | Long Term (per month) Allan Deviation (1 s) Allan Deviation (10 s) Allan Deviation (100 s) |
| Frequency Measure | 1 Hz to 50 GHz | 5 parts in 10^{10} | |
| NVLAP Code: 20/F03 Oscillator Characterization AM Measure (carrier F) | (Rate/Depth) | | |
| 100 kHz to 10 MHz | 50 Hz to 10 kHz/ 5% to 99% | 0.75 % | |
| 10 MHz to 3000 MHz | 50 Hz to 100 kHz/ 20% to 99% | 0.5 % | |
| 10 MHz to 3000 MHz | 50 Hz to 100 kHz/ 5% to 20% | 2.5 % | |
| 3 GHz to 26.5 GHz | 50 Hz to 100 kHz/ 20% to 99% | 1.5 % | |
| 3 GHz to 26.5 GHz | 50 Hz to 100 kHz/ 5% to 20% | 4.5 % | |
| 26.5 GHz to 31.15 GHz | 50 Hz to 100 kHz/ 20% to 99% | 1.9 % | |
| 26.5 GHz to 31.15 GHz | 50 Hz to 100 kHz/ 5% to 20% | 6.8 % | |
| 31.15 GHz to 50 GHz | 50 Hz to 100 kHz/ 20% to 99% | 6 % | |
| 31.15 GHz to 50 GHz | 50 Hz to 100 kHz/ 5% to 20% | 26 % | |
| FM Measure (Carrier F) | (Rate/Peak Deviation) | | |
| 250 kHz to 10 MHz | 20 Hz to 10 kHz/ 20 Hz to 10kHz | $\beta > 0.2$, 1.5 % $\beta > 1.2$, 1.0 % | β is the ratio of the frequency deviation to modulation rate (deviation/rate) |
| 10 MHz to 6.6 GHz | 50 Hz to 200kHz / 250 Hz to 400 kHz | $\beta > 0.2$, 1.5 % $\beta > 0.45$, 1.0 % | |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|----------------------------------------------------|---------------------------------------------------------------------------------------------------|-----------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 6.6 GHz to 13.2 GHz | 50 Hz to 200kHz/ 250 Hz to 400 kHz | $\beta > 0.2$, 2.5 % $\beta > 8$, 1.0 % | The upper limit of the deviation range is determined by a combination of carrier frequency and bandwidth settings. The maximum measured value is 24,999 radians. |
| 13.2 GHz to 31.15 GHz | 50 Hz to 200kHz/ 250 Hz to 400 kHz | $\beta > 0.2$, 3.8 % $\beta > 16$, 1.0 % | |
| 31.15 GHz to 50 GHz | 50 Hz to 200kHz/ 250 Hz to 400 kHz | $\beta > 0.2$, 8.5 % $\beta > 32$, 1.0 % | |
| PM Measure (Carrier F) | (Deviation in Radians) | | |
| 10 MHz to 6.6 GHz | >0.7 | 1.0 % | |
| 6.6 GHz to 13.2 GHz | >0.3 | 3.0 % | |
| 13.2 GHz to 26.5 GHz | >2.0 | 1.0 % | |
| 26.5 GHz to 31.5 GHz | >0.6 | 3.0 % | |
| 31.5 GHz to 50 GHz | >4.0 | 1.0 % | |
| | >1.2 | 3.0 % | |
| | >4.0 | 1.0 % | |
| | >1.3 | 3.0 % | |
| | >8.0 | 1.0 % | |
| | >2.4 | 3.0 % | |
| Total Harmonic Distortion 20 Hz to 250 kHz | -80 dB to 0 dB (0.01 % to 100 %) | 1 dB | |
| Spectral Purity 10 MHz to 18 GHz | 0 dBc to 138 dBc | 1.5 dBc | |
| MECHANICAL | | | |
| NVLAP Code: 20/M06 FORCE Force | 0.2 lbf to 100 lbf 2 lbf to 1000 lbf 1000 lbf to 50 000 lbf 100 000 lbf to 1 000 000 lbf | 0.05 % 0.05 % 0.05 % 0.05 % | Deadweight ASTM-E74 |
| NVLAP Code: 20/M08 MASS Metric | 454 kg 60 kg 50 kg 25 kg 20 kg | 0.71 g 59 mg 49 mg 30 mg 19 mg | Frazier RU100 Mettler-Toledo XP643L |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|---------------------------------------------------------------------------------------------------------|------------------|-----------------------------------------|--------------------------------------------------------------------------------------------------------------|
| | 10 kg | 4 mg | Sartorius CC1000S |
| | 5 kg | 1.9 mg | |
| | 3 kg | 1.2 mg | |
| | 2 kg | 1 mg | |
| | 1 kg | 0.082 mg | Mettler Toledo AT1005 |
| | 500 g | 0.073 mg | |
| | 300 g | 0.077 mg | |
| | 200 g | 0.065 mg | |
| | 100 g | 0.057 mg | Mettler Toledo AX106 |
| | 50 g | 0.067 mg | |
| | 30 g | 0.075 mg | |
| | 20 g | 0.0088 mg | |
| | 10 g | 0.0088 mg | |
| | 5 g | 0.0088 mg | Mettler Toledo XP6U |
| | 3 g | 0.0088 mg | |
| | 2 g | 0.0011 mg | |
| | 1 g | 0.0012 mg | |
| | 500 mg | 0.0012 mg | |
| | 300 mg | 0.00065 mg | |
| | 200 mg | 0.00065 mg | |
| | 100 mg | 0.00065 mg | |
| | 50 mg | 0.00063 mg | |
| | 30 mg | 0.00063 mg | |
| | 20 mg | 0.00063 mg | |
| | 10 mg | 0.00063 mg | |
| | 5 mg | 0.00063 mg | |
| | 3 mg | 0.00063 mg | |
| | 2 mg | 0.00063 mg | |
| | 1 mg | 0.00063 mg | |
| ELECTROMAGNETICS – RF/MICROWAVE | | | |
| NVLAP Code: 20/R04 ELECTROMAGNETIC FIELD STRENGTH Field Strength Probe Calibration | 10 kHz to 40 GHz | 5 % | Uncertainty may increase to 40 % depending on frequency, device under test (DUT), and measurement conditions |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|------------------------------------------------------------------------------------------|----------------|-----------------------------------------------------------|------------------------|
| 0.045 GHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 18 GHz | | 0.002 / 0.4° 0.0015 / 1.1° 0.0015 / 4.0° | VNA & 7 mm TRL Cal Kit |
| 0.045 GHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 26.5 Hz | | 0.004 / 0.63° 0.005 / 2.2° 0.006 / 3.0° | VNA & 3.5 mm Cal Kit |
| 0.045 GHz to 2 GHz 2 GHz to 20 GHz 20 GHz to 40 GHz 40 GHz to 50 GHz | | 0.008 / 0.7° 0.008 / 4° 0.012 / 5.8° 0.015 / 11° | VNA & 2.4 mm Cal Kit |
| Transmission S_{12}/S_{21} 300 kHz to 1.3 GHz 1.3 GHz to 3 GHz 3 GHz to 6 GHz | 0 dB to 20 dB | 0.081 dB 0.089 dB 0.12 dB | VNA & 7 mm Cal Kit |
| 300 kHz to 1.3 GHz 1.3 GHz to 3 GHz 3 GHz to 6 GHz | 20 dB to 50 dB | 0.097 dB 0.097 dB 0.13 dB | |
| 300 kHz to 1.3 GHz 1.3 GHz to 3 GHz 3 GHz to 6 GHz | 50 dB to 60 dB | 0.13 dB 0.15 dB 0.23 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 18 GHz | 0 dB to 20 dB | 0.027 dB 0.023 dB 0.023 dB | VNA & 7 mm TRL Cal Kit |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 18 GHz | 20 dB to 40 dB | 0.037 dB 0.040 dB 0.054 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 18 GHz | 40 dB to 50 dB | 0.075 dB 0.10 dB 0.14 dB | |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|----------------------------------------------------------------------------|----------------|----------------------------------------------|----------------------|
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 18 GHz | 50 dB to 60 dB | 0.21 dB 0.30 dB 0.44 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 20 GHz 20 GHz to 26.5 GHz | 0 dB to 20 dB | 0.044 dB 0.052 dB 0.046 dB 0.054 dB | VNA & 3.5 mm Cal Kit |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 20 GHz 20 GHz to 26.5 GHz | 20 dB to 40 dB | 0.142 dB 0.058 dB 0.057 dB 0.094 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 20 GHz 20 GHz to 26.5 GHz | 40 dB to 50 dB | 0.416 dB 0.079 dB 0.105 dB 0.22 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 20 GHz 20 GHz to 26.5 GHz | 50 dB to 60 dB | 1.35 dB 0.157 dB 0.273 dB 0.651 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 20 GHz 20 GHz to 26.5 GHz | 0 dB to 20 dB | 0.038 dB 0.072 dB 0.15 dB 0.24 dB | VNA & 2.4 mm Cal Kit |
| 45 MHz to 2 GHz 2 GHz to 8 GHz 8 GHz to 20 GHz 20 GHz to 26.5 GHz | 20 dB to 40 dB | 0.14 dB 0.086 dB 0.31 dB 1.0 dB | |
| 45 MHz to 2 GHz 2 GHz to 8 GHz | 40 dB to 50 dB | 0.42 dB 0.14 dB | |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|----------------------------------------------------------------|---------------------|-----------------------------------------|---------|
| 8 GHz to 20 GHz | | 0.76 dB | |
| 20 GHz to 26.5 GHz | | 3.2 dB | |
| 45 MHz to 2 GHz | 50 dB to 60 dB | 1.3 dB | |
| 2 GHz to 8 GHz | | 0.33 dB | |
| 8 GHz to 20 GHz | | 2.3 dB | |
| Attenuation Measure | | | |
| 1 kHz | 0 dB to 60 dB | 0.02 dB | |
| | 60 dB to 110 dB | 0.02 dB + 0.02 dB/10 dB step | |
| 1 KHz to 50 GHz | 0 dB to 10 dB | 0.02 dB | |
| | 10 dB to 30 dB | 0.03 dB | |
| | 30 dB to 50 dB | 0.04 dB | |
| | 50 dB to 70 dB | 0.05 dB | |
| | 70 dB to 90 dB | 0.06 dB | |
| | 90 dB to 110 dB | 0.07 dB | |
| | 110 dB to 120 dB | 0.08 dB | |
| Relative Tuned Power | | | |
| 2.5 MHz to 50 GHz | 10 dB to 0 dB | 0.03 dB | |
| | 0 dB to -10 dB | 0.02 dB | |
| | -10 dB to -30 dB | 0.03 dB | |
| | -30 dB to -50 dB | 0.04 dB | |
| | -50 dB to -70 dB | 0.05 dB | |
| | -70 dB to -90 dB | 0.06 dB | |
| | -90 dB to -110 dB | 0.07 dB | |
| | -110 dB to -120 dB | 0.08 dB | |
| NVLAP Code: 20/R17 RF/ Microwave Power Meters | | | |
| Absolute Tuned Power | | | |
| 2.5 MHz to 50 GHz | 10 dBm to 0 dBm | 0.17 dB | |
| | 0 dBm to -40 dBm | 0.18 dB | |
| | -40 dBm to -70 dBm | 0.22 dB | |
| | -70 dBm to -90 dBm | 0.24 dB | |
| | -90 dBm to -110 dBm | 0.28 dB | |

2012-01-01 through 2012-12-31

Effective dates

David E. Alderman

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty (<i>k</i> =2) ^{Note 3} | Remarks |
|-------------------------------------------------------------|-----------------------------------------------------------|----------------------------------------------|-----------------------------------------------|
| Power Sensor Cal Factor | -110 dBm to -120 dBm | 0.31 dB | |
| 100 kHz to 10 GHz | 10 μW to 100 mW | 1.6 % | |
| 10 GHz to 18 GHz | 10 μW to 100 mW | 1.8 % | |
| 18 GHz to 26.5 GHz | 10 μW to 100 mW | 2.5 % | |
| 26.5 GHz to 40 GHz | 10 μW to 100 mW | 3.2 % | |
| 40 GHz to 50 GHz | 10 μW to 100 mW | 3.8 % | |
| 1 GHz to 18 GHz | 100 pW to 10 μW | 2.2 % | |
| 18 GHz to 26.5 GHz | 100 pW to 10 μW | 3.7 % | |
| THERMODYNAMICS | | | |
| NVLAP Code: 20/T02 Humidity | | | |
| Relative Humidity (RH) | 8 % RH to 95 % RH | 1.0 % RH | -20° C to 60° C |
| Dew Point | -70° C to -35° C -35° C to 20° C | 0.30° C 0.25° C | Sensor at 17° C to 23° C |
| NVLAP Code: 20/T03 Laboratory Thermometers | | | |
| | -100° C to 0° C | 0.040° C | SPRT using Liquid Bath |
| | 0° C to 30° C | 0.02° C | |
| | 30° C to 300° C | 0.03° C | |
| | 300° C to 1000° C | 1.0° C | Dry Well Furnace with S Type TC |
| NVLAP Code: 20/T05 Pressure | | | |
| Pneumatic Generate | 1 psi to 25 psi 5 psi to 114 psi 10 psi to 1000 psi | 0.0035 % 0.0035 % 0.0040 % | Ruska 2465 Piston Gage with Dead Weights |
| Hydraulic Generate | 100 psi to 20 000 psi 1 500 to 72 500 psi | 0.0060 % 0.012 % | Ruska 2485 & DH Piston Gage with Dead Weights |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

| Measured Parameter or Device Calibrated | Range | Uncertainty ($k=2$) ^{Note 3} | Remarks |
|-----------------------------------------|---------------------|--------------------------------------------|------------------------------------|
| <i>NVLAP Code: 20/T09</i> | (in inches of Hg) | Absolute/Differential (in inches of Hg) | |
| Vacuum & Low Pressure Gages | 0 | NA / 0.00009 | Schwien 110 inch Mercury Manometer |
| | 0.1 | 0.000171 / 0.000091 | |
| | 0.5 | 0.000174 / 0.000091 | |
| | 1 | 0.000178 / 0.000091 | |
| | 2 | 0.000185 / 0.000092 | |
| | 5 | 0.000207 / 0.000095 | |
| | 10 | 0.000245 / 0.0001 | |
| | 20 | 0.00032 / 0.000109 | |
| | 30 | 0.000394 / 0.000118 | |
| | 40 | 0.000469 / 0.000127 | |
| | 60 | 0.000619 / 0.000145 | |
| 90 | 0.000843 / 0.000172 | | |
| 110 | 0.000993 / 0.00019 | | |
| END | | | |

2012-01-01 through 2012-12-31

Effective dates

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200762-0

Scope Revised: 2012-01-06

Notes

Note 1: A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

Note 2: Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

Note 3: The uncertainty associated with a measurement in a CMC is an expanded uncertainty using a coverage factor, k = 2, with a level of confidence of approximately 95 %. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

Note 3a: The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

Note 3b: As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

Note 3c: As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under normal conditions. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.1.h. of NIST Handbook 150, Procedures and General Requirements.

Note 4: Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

Note 5: Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

Note 6: NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

Note 7: See NIST Handbook 150 for further explanation of these notes.

2012-01-01 through 2012-12-31

Effective dates

David Z. Alderman

For the National Institute of Standards and Technology