

National Institute
of Standards and Technology



National Voluntary
Laboratory Accreditation Program

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Page 1 of 1

CALIBRATION LABORATORIES

NVLAP LAB CODE 200493-0

GRETAGMACBETH LLC STANDARDS LABORATORY

617 Little Britian Road
New Windsor, NY 12553-2511
Mr. Harold R. Van Aken
Phone: 845-565-7660 x515 Fax: 845-561-8140
E-Mail: vanaken@gretagmacbeth.com

OPTICAL RADIATION

NVLAP Code: 20/O02
Color Temperature

| <i>Range in °kelvin</i> | <i>Best Uncertainty (±) in °kelvin^{note 1}</i> | <i>Remarks</i> |
|-------------------------|---|----------------|
| 2300 | 24 | |
| 2856 | 20 | |

1. Represents an expanded uncertainty using a coverage factor, $k=2$.

June 30, 2005

Effective through

A handwritten signature in black ink, appearing to read 'Harold R. Van Aken', written over a horizontal line.

For the National Institute of Standards and Technology