

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 5/13/2004

Page 1 of 1

CALIBRATION LABORATORIES

NVLAP LAB CODE 200235-0

STERIS-ISOMEDIX SERVICES

7828 Nagle Avenue
Morton Grove, IL 60053
Mr. Deepak Patil
Phone: 847-966-1160 Fax: 847-965-2855
E-Mail: deepak_patil@steris.com

IONIZING RADIATION

NVLAP Code: 20/I02
High Dose Dosimetry

Source

⁶⁰Co Gamma Rays

Range

100 grays or more at a rate of
approximately 1 to 20 kGy/hour

Best Uncertainty(±)^{note 1}

2.5%

1. Represents an expanded uncertainty using a coverage factor, k=2.

December 31, 2004

Effective through

For the National Institute of Standards and Technology