

ISO/IEC 17025:1999  
ISO 9002:1994

## Scope of Accreditation



Revised 5/20/2004

Page 1 of 1

CALIBRATION LABORATORIES

NVLAP LAB CODE 200536-0

ASANUMA GIKEN CO., LTD.  
4079-1 Koto-Cho, Hamamatsu-shi  
Shizuoka-ken 431-1103  
JAPAN  
Mr. Susumu Asanuma  
Phone: 81-53-486-1110 Fax: 81-53-486-1109  
E-Mail: papa@asanuma-g.jp

### DIMENSIONAL

*NVLAP Code:* 20/D16  
Coordinate Measuring Machines

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks<sup>note 2</sup></i>
0 to 150 mm	3.0 $\mu$ m	Artifact Calibration

1. Represents an expanded uncertainty using a coverage factor,  $k=2$ .
2. Calibration, including two and three dimensional measurements of the Asanuma Giken manufactured Quality Master Gauge.

March 31, 2005

Effective through

For the National Institute of Standards and Technology